

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

North Carolina

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



2014

Scope

Mass Echelon III

1000 kg to 1 mg
2500 lb to 0.001 lb
4 oz to 0.03125 oz

Weight Carts
6000 lb to 3000 lb

Lottery Balls

Volume Gravimetric, I

400 L to 100 mL
100 gal to 1 gill
10 ft³ to 0.5 ft³

Volume Gravimetric, SVP

30 gal to 15 gal

Volume Transfer, II

650 L to 20 L
2000 gal to 1 gal
500 gal to 25 gal LPG

Length

Lottery Balls

A handwritten signature in blue ink, appearing to read "Carol T. Hockert", is written over a horizontal line.

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2014-07-15 to 2014-12-31

Amended: 2014-07-15

Expanded Volume Gravimetric
and Transfer Scope.

NOTE: This Scope does not cover
calibrations at remote locations.