



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

<p>North Carolina Standards Laboratory 4040 District Drive / 1051 Mail Service Center Raleigh, NC 27699-1051 Ms. Sharon Woodard Phone: 919-733-4411 x213 Fax: 919-733-8804 E-mail: sharon.woodard@ncagr.gov URL: http://www.ncagr.gov/standard</p>	<p>Fields of Calibration Dimensional Mechanical</p> <p>This laboratory is compliant to ANSI/NCSL Z540-1-1994; Part 1. (20/A01)</p>
---	---

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
DIMENSIONAL			
LENGTH and DIAMETER, STEP GAGES (20/D05)			
Lottery Ball Pass Through Gauge	Lottery Ball	7.5 μ m	
MECHANICAL			
MASS DETERMINATION (20/M08)			
Metric	30 kg 20 kg 10 kg 5 kg 3 kg 2 kg 1 kg 500 g 300 g 200 g 100 g 50 g 30 g 20 g 10 g 5 g 3 g 2 g 1 g	8.0 mg 5.7 mg 4.0 mg 0.48 mg 0.30 mg 0.22 mg 65 μ g 33 μ g 21 μ g 14 μ g 12 μ g 6.4 μ g 4.5 μ g 3.7 μ g 3.5 μ g 1.9 μ g 1.4 μ g 1.1 μ g 1.0 μ g	Echelon I

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) <small>Note 3</small>	Remarks
Metric	500 mg	0.92 μ g	Echelon II
	300 mg	0.78 μ g	
	200 mg	0.76 μ g	
	100 mg	0.80 μ g	
	50 mg	0.52 μ g	
	30 mg	0.38 μ g	
	20 mg	0.32 μ g	
	10 mg	0.30 μ g	
	5 mg	0.20 μ g	
	3 mg	0.14 μ g	
	2 mg	0.12 μ g	
	1 mg	0.12 μ g	
	30 kg	8.0 mg	
	20 kg	5.7 mg	
	10 kg	4.0 mg	
	5 kg	0.48 mg	
	3 kg	0.30 mg	
	2 kg	0.22 mg	
	1 kg	65 μ g	
	500 g	33 μ g	
	300 g	21 μ g	
	200 g	14 μ g	
	100 g	12 μ g	
	50 g	6.4 μ g	
	30 g	4.5 μ g	
	20 g	3.7 μ g	
	10 g	3.5 μ g	
	5 g	1.9 μ g	
	3 g	1.4 μ g	
	2 g	1.1 μ g	
	1 g	1.0 μ g	
	500 mg	0.92 μ g	
	300 mg	0.78 μ g	
200 mg	0.76 μ g		
100 mg	0.80 μ g		

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
Metric	50 mg	0.52 µg	Echelon III
	30 mg	0.38 µg	
	20 mg	0.32 µg	
	10 mg	0.30 µg	
	5 mg	0.20 µg	
	3 mg	0.14 µg	
	2 mg	0.12 µg	
	1 mg	0.12 µg	
	1000 kg	17 g	
	500 kg	3.5 g	
	250 kg	1.9 g	
	200 kg	1.9 g	
	100 kg	1.6 g	
	50 kg	0.30 g	
	30 kg	0.13 g	
	25 kg	0.13 g	
	20 kg	0.13 g	
	10 kg	76 mg	
	5 kg	30 mg	
	3 kg	12 mg	
	2 kg	12 mg	
	1 kg	6.7 mg	
	500 g	4.3 mg	
	300 g	3.3 mg	
	200 g	1.2 mg	
	100 g	0.60 mg	
	50 g	0.37 mg	
	30 g	0.31 mg	
	20 g	0.31 mg	
	10 g	0.26 mg	
	5 g	0.18 mg	
	3 g	0.11 mg	
2 g	89 µg		
1 g	60 µg		
500 mg	45 µg		

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) Note 3	Remarks
	300 mg	36 μ g	
	200 mg	31 μ g	
	100 mg	24 μ g	
	50 mg	19 μ g	
	30 mg	17 μ g	
	20 mg	15 μ g	
	10 mg	12 μ g	
	5 mg	10 μ g	
	3 mg	9.0 μ g	
	2 mg	7.9 μ g	
	1 mg	6.9 μ g	
Lottery Balls	Lottery Ball	43 mg	Echelon III
Avoirdupois	2500 lb	21 g	Echelon III
	2000 lb	16 g	
	1000 lb	2.8 g	
	500 lb	1.9 g	
	100 lb	0.60 g	
	50 lb	0.13 g	
	30 lb	76 mg	
	25 lb	56 mg	
	20 lb	55 mg	
	10 lb	18 mg	
	5 lb	12 mg	
	4 lb	6.8 mg	
	3 lb	6.8 mg	
	2 lb	4.7 mg	
	1 lb	3.9 mg	
	0.5 lb	3.6 mg	
	0.3 lb	1.1 mg	
	0.2 lb	0.67 mg	
	0.1 lb	0.49 mg	
	0.05 lb	0.37 mg	
	0.03 lb	0.26 mg	
	0.02 lb	0.18 mg	

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) ^{Note 3}	Remarks
Weight Carts	0.01 lb	0.15 mg	Echelon III
	0.005 lb	0.13 mg	
	0.003 lb	61 µg	
	0.002 lb	46 µg	
	0.001 lb	36 µg	
	4 oz	1.1 mg	
	2 oz	0.67 mg	
	1 oz	0.37 mg	
	1/2 oz	0.26 mg	
	1/4 oz	0.18 mg	
	1/8 oz	0.15 mg	
	1/16 oz	60 µg	
	1/32 oz	46 µg	
	6000 lb	120 g	
	5500 lb	100 g	
	5000 lb	100 g	
	4500 lb	83 g	
	4000 lb	83 g	
3000 lb	71 g		
VOLUME and DENSITY (20/M12)			
Volume	2000 gal	81 in ³	Transfer Method 4 in neck
	1500 gal	62 in ³	
	1200 gal	48 in ³	
	1000 gal	42 in ³	
	500 gal	21 in ³	
	200 gal	2.8 in ³	
	100 gal	1.9 in ³	
	60 gal	1.4 in ³	
	50 gal	1.1 in ³	
	25 gal	0.73 in ³	
	15 gal	0.52 in ³	
	10 gal	0.34 in ³	
	5 gal	0.21 in ³	

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) <small>Note 3</small>	Remarks
Test Measure	5 gal	0.19 in ³	3 in neck
	1 gal	0.12 in ³	
	650 L	2.5 in ³	
	500 L	2.3 in ³	
	40 L	0.35 in ³	
	20 L	0.22 in ³	LPG Transfer Method
	500 gal	30 in ³	
	100 gal	3.8 in ³	
	50 gal	4.1 in ³	
	25 gal	3.2 in ³	Gravimetric Method
	100 gal	0.70 in ³	
	75 gal	2.8 in ³	
	50 gal	0.67 in ³	
	25 gal	0.53 in ³	
	15 gal	0.33 in ³	4 in neck 3 in neck 2 in neck
5 gal	0.30 in ³		
5 gal	0.25 in ³		
5 gal	0.11 in ³		
1 gal	0.24 in ³		
10 ft ³	1.5 in ³		
0.5 ft ³	0.062 in ³	3.9 mL	
20 L	3.9 mL		
Prover	400 L		31 mL
Prover	250 L	100 mL	2.7 mL
	Graduated Cylinder	2 L	
Flask	100 mL	0.049 mL	0.39 mL 0.065 mL
	1 qt	0.39 mL	
	1 gill	0.065 mL	

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) Notes 1,2

Measured Parameter or Device Calibrated	Range	Uncertainty ($k=2$) <small>Note 3</small>	Remarks
Slicker Standard	100 gal	0.75 in ³	
	50 gal	0.50 in ³	
	25 gal	0.17 in ³	
	15 gal	0.17 in ³	
	5 gal	0.040 in ³	
	1 gal	0.021 in ³	
	1.0 ft ³	0.17 in ³	
	20 L	1.1 mL	
	10 L	1.6 mL	
	5 L	0.83 mL	
Small Volume Prover	30 gal	1.2 in ³	Gravimetric Method
	20 gal	0.66 in ³	
	15 gal	0.78 in ³	
END			

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology



CALIBRATION LABORATORIES

NVLAP LAB CODE 200495-0

Scope Revised: 2015-01-26

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of $k = 2$. However, laboratories may report a coverage factor different than $k = 2$ to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.1.h. of NIST Handbook 150, Procedures and General Requirements.

Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

Note 7: See NIST Handbook 150 for further explanation of these notes

2014-04-01 through 2015-03-31

Effective dates

For the National Institute of Standards and Technology