Animal Welfare Section
NC Department of Agriculture and Consumer Services
1030 Mail Service Center
Raleigh, NC 27699-1030
phone: (919) 715-7111 FAX: (919) 733-6431
e-mail: agr.aws@ncagr.gov URL: www.ncaws.com

Animal Welfare Section, NCDA&CS
Euthanasia Inspection Report

Name of business: Sampson County Animal Control + Pet Adoption

City: Clinton | License number (if currently licensed): 81 | license type: 44

Duties of CETs

Prepare animals for euthanasia .0418: Acceptable	Properly record all data .0418: Acceptable	Security, controlled substances .0418: Acceptable
Supervise Prob. CET .0418: N/A	Properly euthanize .0418: Acceptable	Properly dispose of dead .0418: Acceptable

Euthanasia by injection

IC only on anesth. or sedated .0501: N/A

Euthanasia by CO

Use only bottled gas .0601: Acceptable	Use only comm. mfd chamber .0601: Acceptable	Only same species in chamber .0601: Acceptable	In chamber for >= 20 min. .0601: Acceptable
Not used on < 16 weeks .0602: Acceptable	Not used on pregnant .0602: Acceptable	Not used on near death .0602: Acceptable	No live with dead .0603: Acceptable
Animals separated .0604: Acceptable	At least 1 viewport .0605: Acceptable	Chamber in good order .0605: Acceptable	Airtight seals present .0605: Acceptable
Light shatterproof .0605: N/A	Chamber sufficiently lit .0605: Acceptable	Electrical explosion-proof .0605: Acceptable	If inside, two CO monitors .0605: N/A

Records of monthly inspection .0606: Acceptable	Records of yearly inspection .0606: Acceptable	Visual inspection by AWS: Acceptable

Chamber cleaned b/t uses .0607: Acceptable	Operational guide & or manual .0608: Acceptable	>= 2 adults present when used .0609: Acceptable

Extraordinary methods

Reports of extraordinary euth. .0705: N/A

Policy and procedure manual

Current copy of AWA in manual .0803: Acceptable	Current AVMA euth. in manual .0803: Acceptable	Current HSUS euth. in manual .0803: Acceptable	Current AHA euth. in manual .0803: Acceptable
List of approved euth. methods .0803: Acceptable	List of CETs & methods .0803: Acceptable	Contact info for DVM in PVC .0803: Acceptable	Contact info for DVM care . 0803: Acceptable
List after hour euth. meth. 0803: Acceptable	Euth. methods if no CET present 0803: Acceptable	Policy for verifying death . 0803: Acceptable	Contact info for suppliers. 0803: Acceptable
DEA certificate . 0803: N/A	MSDS sheets, chemical or gas . 0803: Acceptable	MSDS sheets, tranq. or anesth. . 0803: Acceptable	Signs & symptoms, human . 0803: Acceptable
First aid information . 0803: Acceptable	MD contact information . 0803: Acceptable		

Pam. Shelar — Signature of inspector
10/8/09 — date
page 1 of 2
Susan J Holder — Signature of management

Euth. Insp.
Cont.

NCDA&CS, VETERINARY DIVISION
ANIMAL WELFARE SECTION,
1030 MAIL SERVICE CENTER,
RALEIGH, NC 27699-1030
PHONE: 919/715-7111, FAX: 919/733-6431

ANIMAL WELFARE INSPECTION CONTINUATION PAGE

LICENSE #: 81
TYPE FACILITY: Animal Shelter (Private/Public) ☒ Boarding Kennel ☐ Pet Shop ☐ Public Auction ☐
BUSINESS NAME: Sampson Co. AC + Pet Adoption
OWNER:
ADDRESS:
TELEPHONE: (_____)_____-_____

Cont.

Item Number	Explanation of Inadequacy (circled items above) And Recommendation For Compliance	Date Corrections Must Be Completed
Follow-up from 9/25/09		
Comments:		
All unacceptable items have [redacted] been addressed.		
There are no sanitation issues around the chamber today.		
The seals have been replaced and are airtight.		
The chamber is in good order.		
Monthly records of inspection are available today.		
The yearly inspection was performed 10/7/09 and the record was available.		

☒ APPROVED ☐ CONDITIONALLY APPROVED ☐ DISAPPROVED Date: 10/8/09 Time: 3:00pm

R/M. Shelar
Inspector's Signature

Susan Holder
Owner/Authorized Agent's Signature

AW-2
Rev. 1/07

White= Office Canary= Inspector Pink= Owner

FROM: RALPH W. HOUSER DVM
CAROLINA VETERINARY CONSULTING
10020 S. NC87
PITTSBORO, NC 27312

TO: SUSAN HOLDER, DIRECTOR
SAMPSON CO. ANIMAL SHELTER
168 AGRICULTURE PLACE
CLINTON, NC 28328

RE: CHAMBER INSPECTION 07 OCTOBER 2009

CHECKING ALL PIPING, DOOR-SEALS, VIEW PORTS, AND REGULATOR WITH A CALIBRATED MSA MINICO METER, NO CO READINGS WERE ABOVE BACKGROUND. THIS INDICATES NO CO LEAKS

AT 20 PSI FOR 25 SECONDS, THIS SYSTEM DELIVERS 10% CO IN THE CHAMBER AS DESIGNED.

AFTER RUNNING THE VENTILATION SYSTEM FOR 10 MINUTES, THE CO READINGS IN THE CHAMBER WERE THE SAME AS BACKGROUND READINGS OUTSIDE THE CHAMBER.

THIS CHAMBER IS OPERATING AS DESIGNED AND IS SAFE FOR PERSONNEL.